



# UNITED STATES PATENT AND TRADEMARK OFFICE

UNITED STATES DEPARTMENT OF COMMERCE  
United States Patent and Trademark Office  
Address: COMMISSIONER FOR PATENTS  
P.O. Box 1450  
Alexandria, Virginia 22313-1450  
www.uspto.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
10/561,515	12/20/2005	Owen Shawn Sobelman	17970USPCT (HEA)	9886
51957	7590	06/22/2009	EXAMINER	
ALLERGAN, INC. 2525 DUPONT DRIVE, T2-7H IRVINE, CA 92612-1599			TRUONG, KEVIN THAO	
			ART UNIT	PAPER NUMBER
			3734	
			MAIL DATE	DELIVERY MODE
			06/22/2009	PAPER

**Please find below and/or attached an Office communication concerning this application or proceeding.**

The time period for reply, if any, is set in the attached communication.

<b>Office Action Summary</b>	<b>Application No.</b> 10/561,515	<b>Applicant(s)</b> SOBELMAN ET AL.	
	<b>Examiner</b> Kevin T. Truong	<b>Art Unit</b> 3734	

-- The MAILING DATE of this communication appears on the cover sheet with the correspondence address --

### Period for Reply

A SHORTENED STATUTORY PERIOD FOR REPLY IS SET TO EXPIRE 3 MONTH(S) OR THIRTY (30) DAYS, WHICHEVER IS LONGER, FROM THE MAILING DATE OF THIS COMMUNICATION.

- Extensions of time may be available under the provisions of 37 CFR 1.136(a). In no event, however, may a reply be timely filed after SIX (6) MONTHS from the mailing date of this communication.
- If NO period for reply is specified above, the maximum statutory period will apply and will expire SIX (6) MONTHS from the mailing date of this communication.
- Failure to reply within the set or extended period for reply will, by statute, cause the application to become ABANDONED (35 U.S.C. § 133). Any reply received by the Office later than three months after the mailing date of this communication, even if timely filed, may reduce any earned patent term adjustment. See 37 CFR 1.704(b).

### Status

- 1) ☐ Responsive to communication(s) filed on \_\_\_\_.
- 2a) ☐ This action is **FINAL**. 2b) ☒ This action is non-final.
- 3) ☐ Since this application is in condition for allowance except for formal matters, prosecution as to the merits is closed in accordance with the practice under *Ex parte Quayle*, 1935 C.D. 11, 453 O.G. 213.

### Disposition of Claims

- 4) ☒ Claim(s) 1-35 is/are pending in the application.
- 4a) Of the above claim(s) \_\_\_\_ is/are withdrawn from consideration.
- 5) ☐ Claim(s) \_\_\_\_ is/are allowed.
- 6) ☒ Claim(s) 1-35 is/are rejected.
- 7) ☐ Claim(s) \_\_\_\_ is/are objected to.
- 8) ☐ Claim(s) \_\_\_\_ are subject to restriction and/or election requirement.

### Application Papers

- 9) ☐ The specification is objected to by the Examiner.
- 10) ☐ The drawing(s) filed on \_\_\_\_ is/are: a) ☐ accepted or b) ☐ objected to by the Examiner.  
Applicant may not request that any objection to the drawing(s) be held in abeyance. See 37 CFR 1.85(a).  
Replacement drawing sheet(s) including the correction is required if the drawing(s) is objected to. See 37 CFR 1.121(d).
- 11) ☐ The oath or declaration is objected to by the Examiner. Note the attached Office Action or form PTO-152.

### Priority under 35 U.S.C. § 119

- 12) ☐ Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).
- a) ☐ All b) ☐ Some \* c) ☐ None of:
1. ☐ Certified copies of the priority documents have been received.
  2. ☐ Certified copies of the priority documents have been received in Application No. \_\_\_\_.
  3. ☐ Copies of the certified copies of the priority documents have been received in this National Stage application from the International Bureau (PCT Rule 17.2(a)).

\* See the attached detailed Office action for a list of the certified copies not received.

### Attachment(s)

- |  |   |
|--|---|
| 1) <input checked="" type="checkbox"/> Notice of References Cited (PTO-892)            | 4) <input type="checkbox"/> Interview Summary (PTO-413)           |
| 2) <input type="checkbox"/> Notice of Draftsperson's Patent Drawing Review (PTO-948)   | Paper No(s)/Mail Date. ____.                                      |
| 3) <input checked="" type="checkbox"/> Information Disclosure Statement(s) (PTO/SB/08) | 5) <input type="checkbox"/> Notice of Informal Patent Application |
| Paper No(s)/Mail Date <u>12/05; 5/08</u> .   | 6) <input type="checkbox"/> Other: ____.                          |

## DETAILED ACTION

### ***Claim Rejections - 35 USC § 102***

1. The following is a quotation of the appropriate paragraphs of 35 U.S.C. 102 that form the basis for the rejections under this section made in this Office action:

A person shall be entitled to a patent unless –

(b) the invention was patented or described in a printed publication in this or a foreign country or in public use or on sale in this country, more than one year prior to the date of application for patent in the United States.

2. Claims 1-23 and 34-35 are rejected under 35 U.S.C. 102(b) as being anticipated by Akui et al. (US 4715360).

Note in figures 2 and 11, a valve body (32,72) having proximal and distal ends; a flange surface (50) disposed at the proximal end of the valve body (32,72); a first chamber (at 53) formed in said valve body (32) for accepting an inflation tube (26) inserted through said opening in said flange surface (50); a concave section (portion below (54)) located adjacent said first chamber (at 53); and a slit (64) formed in the valve body (72) connecting said concave section (portion below (54)); a second chamber (at 54) located between said first chamber (53) and said concave section (portion below (54)); wherein said flange surface opening can be accepted forceps (26) rather a inflation tube; wherein upon insertion of the inflation tube (26) engages a neck portion (58,66); wherein pressurized fluid flowing through the inflation tube (26) causes said slit (64) to open allowing fluid to pass therethrough and through said distal end of said valve body. A recitation of the intended use of the claimed invention must result in a structural difference between the claimed invention and the prior art in order to

Art Unit: 3734

patentably distinguish the claimed invention from the prior art. If the prior art structure is capable of performing the intended use, then it meets the claim.

***Claim Rejections - 35 USC § 103***

3. The following is a quotation of 35 U.S.C. 103(a) which forms the basis for all obviousness rejections set forth in this Office action:

(a) A patent may not be obtained though the invention is not identically disclosed or described as set forth in section 102 of this title, if the differences between the subject matter sought to be patented and the prior art are such that the subject matter as a whole would have been obvious at the time the invention was made to a person having ordinary skill in the art to which said subject matter pertains. Patentability shall not be negated by the manner in which the invention was made.

4. Claims 1-35 are rejected under 35 U.S.C. 103(a) as being unpatentable over Gau et al. (US 5084061) in view of Akui et al. (US 4715360).

Gau et al discloses the claimed invention except for valve body (28) having concave section located adjacent the first chamber and a slit formed in the valve body connecting the concave section and the distal end of valve body.

Akui et al teaches in figures 2 and 11 that it is know in the art to have the valve body as recited in the claims (as stated above). Therefore, it would have been obvious to one of ordinary skill in the art at the time the invention was made to modify the valve body of Gau et al with the valve body as described by Akui et al in order to prevent leakage of air or filth from the patient's coelomic cavity.

***Conclusion***

5. The prior art made of record and not relied upon is considered pertinent to applicant's disclosure. Nosel et al. (US 6439541) discloses a duckbill valve.

Any inquiry concerning this communication or earlier communications from the examiner should be directed to Kevin T. Truong whose telephone number is 571-272-

Art Unit: 3734

4705. The examiner can normally be reached on Monday-Friday from 8:00 AM to 4:30 PM..

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Todd Manahan can be reached on 571-272-4713. The fax phone number for the organization where this application or proceeding is assigned is 571-273-8300.

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see <http://pair-direct.uspto.gov>. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free). If you would like assistance from a USPTO Customer Service Representative or access to the automated information system, call 800-786-9199 (IN USA OR CANADA) or 571-272-1000.

/Kevin T. Truong/  
Primary Examiner, Art Unit 3734

Kevin T. Truong  
Primary Examiner  
Art Unit 3734